

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3901

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2023 11:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.8 F 0.0  
2:00 190.4 F 3.9  
4:00 207.4 F 8.8  
6:00 221.2 F 7.2  
8:00 230.5 F 9.2  
10:00 239.3 F 9.7  
12:00 242.6 F 12.2  
14:00 241.1 F 10.8  
16:00 254.5 F 17.6  
18:00 266.1 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.7  
2:00 271.9 F 28.7  
3:00 272.2 F 28.6  
4:00 271.9 F 28.7  
  
Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 205.9 F 0.0  
10:00 221.3 F 0.0  
15:00 220.4 F 0.0  
20:00 217.4 F 0.0  
25:00 215.3 F 0.0  
30:00 213.8 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 19:36  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

10/13/2023 12:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_