

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000155  
9:06 15/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:18  
PRESSURIZING 4:18  
STERILIZING 5:55  
136.4 C 225kPa 5:55  
136.3 C 225kPa 6:25  
136.4 C 226kPa 6:55  
136.4 C 225kPa 7:25  
136.2 C 223kPa 7:55  
136.4 C 225kPa 8:25  
136.3 C 224kPa 8:55  
136.2 C 224kPa 9:25  
136.2 C 224kPa 9:25  
VENTING 9:25  
AIR DRYING 10:29  
CYCLE COMPLETE 12:03  
-----

Digital Signature #  
04DB9B1D0DD0710C  
-----