

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000973  
12:24 09/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:07  
PRESSURIZING 2:12  
STERILIZING 3:11  
136.0 C 229kPa 3:11  
136.2 C 232kPa 3:41  
136.4 C 233kPa 4:11  
136.3 C 233kPa 4:41  
136.4 C 233kPa 5:11  
136.4 C 233kPa 5:41  
136.3 C 233kPa 6:11  
136.4 C 232kPa 6:41  
136.4 C 232kPa 6:41  
VENTING 6:41  
AIR DRYING 7:33  
CYCLE COMPLETE 7:42  
-----

Digital Signature #  
85AB3414A401C349  
-----