

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001670  
15:05 03/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:47  
STERILIZING 3:55  
136.4 C 226kPa 3:55  
136.4 C 225kPa 4:25  
136.4 C 226kPa 4:55  
136.3 C 224kPa 5:25  
136.3 C 224kPa 5:55  
136.0 C 222kPa 6:25  
136.4 C 225kPa 6:55  
136.1 C 221kPa 7:25  
136.1 C 221kPa 7:25  
VENTING 7:25  
AIR DRYING 8:12  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
1943E7095C63CD07  
-----