

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004286  
9:12 02/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:40  
STERILIZING 3:46  
136.5 C 221kPa 3:46  
136.4 C 225kPa 4:16  
136.0 C 221kPa 4:46  
136.3 C 219kPa 5:16  
136.1 C 221kPa 5:46  
136.1 C 219kPa 6:16  
136.4 C 224kPa 6:46  
136.3 C 222kPa 7:16  
136.3 C 222kPa 7:16  
VENTING 7:17  
AIR DRYING 8:06  
CYCLE COMPLETE 21:36

-----  
Digital Signature #  
B027FFAB19CDD2D4  
-----