
Midmark M11 - v1.0.0
Total Cycles: 4125

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.2 F 0.0
2:00 120.7 F 1.3
4:00 150.6 F 3.9
6:00 175.8 F 7.1
8:00 192.3 F 10.6
10:00 219.0 F 6.1
12:00 214.6 F 7.3
14:00 226.3 F 11.3
16:00 228.9 F 6.7
18:00 239.1 F 11.8
20:00 237.3 F 9.2
22:00 249.5 F 14.8
24:00 259.8 F 20.6
26:00 269.6 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 28.9
2:00 272.1 F 29.2
3:00 271.4 F 28.6
4:00 272.5 F 29.2

Min 271.1 F 28.5
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 202.6 F 0.0
10:00 191.7 F 0.0
15:00 189.5 F 0.0
20:00 186.4 F 0.0
25:00 185.8 F 0.0
30:00 183.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 27:44
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:06:30
CYCLE COMPLETE

12/15/2023 02:29 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____