

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001008  
12:25 16/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:16  
STERILIZING 3:17  
136.4 C 227kPa 3:17  
136.4 C 233kPa 3:47  
136.4 C 233kPa 4:17  
136.4 C 233kPa 4:47  
136.2 C 232kPa 5:17  
136.4 C 233kPa 5:47  
136.4 C 233kPa 6:17  
136.3 C 232kPa 6:47  
136.3 C 232kPa 6:47  
VENTING 6:47  
AIR DRYING 7:39  
CYCLE COMPLETE 8:57  
-----

Digital Signature #  
DDD20CD6FFD9859E  
-----