
Midmark M11 - v1.0.5
Total Cycles: 955

Name: RITTER2
Sterilizer ID: V2313228
04/05/2022 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.5 F 0.0
2:00 144.4 F 0.0
4:00 150.4 F 0.2
6:00 187.3 F 1.4
8:00 205.2 F 6.2
10:00 223.9 F 10.9
12:00 228.6 F 7.1
14:00 241.2 F 13.1
16:00 241.3 F 11.7
18:00 241.4 F 11.2
20:00 255.3 F 18.4
22:00 267.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.7
3:00 272.6 F 29.5
4:00 271.9 F 28.8

Min 271.4 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 218.8 F 0.0
10:00 217.7 F 0.0
15:00 217.0 F 0.0
20:00 220.6 F 0.0
25:00 227.5 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

04/05/2022 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____