

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001028  
13:41 21/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:06  
STERILIZING 4:07  
136.4 C 220kPa 4:07  
136.4 C 225kPa 4:37  
136.3 C 225kPa 5:07  
136.4 C 223kPa 5:37  
136.3 C 225kPa 6:07  
136.5 C 225kPa 6:37  
136.3 C 223kPa 7:07  
136.4 C 223kPa 7:37  
136.4 C 223kPa 7:37  
VENTING 7:37  
AIR DRYING 8:29  
CYCLE COMPLETE 8:44

-----  
Digital Signature #  
9D6759BE58157948  
-----