
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000228
10:35 09/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 4:00
136.0 C 220kPa 4:00
136.2 C 224kPa 4:30
136.3 C 224kPa 5:00
136.3 C 224kPa 5:30
136.2 C 224kPa 6:00
136.4 C 225kPa 6:30
136.4 C 224kPa 7:00
136.4 C 225kPa 7:30
136.4 C 225kPa 7:30
VENTING 7:30
AIR DRYING 8:19
CYCLE COMPLETE 13:46

Digital Signature #
84988CEFBD364FA0
