

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1055

Name: RITTER2  
Sterilizer ID: V2313228  
05/02/2022 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.8 F 0.0  
2:00 171.4 F 3.4  
4:00 184.2 F 7.7  
6:00 205.9 F 13.0  
8:00 222.8 F 7.6  
10:00 234.5 F 12.9  
12:00 237.0 F 10.7  
14:00 236.4 F 9.4  
16:00 250.7 F 16.3  
18:00 262.5 F 23.3  
20:00 272.5 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 29.9  
2:00 271.6 F 29.2  
3:00 272.0 F 30.0  
4:00 272.5 F 29.7  
  
Min 271.3 F 29.1  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 227.9 F 0.0  
10:00 211.1 F 0.0  
15:00 207.5 F 0.0  
20:00 215.1 F 0.0  
25:00 228.4 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 2:27  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:53  
CYCLE COMPLETE

05/02/2022 12:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_