
Midmark M11 - v1.0.5
Total Cycles: 2801

Name: RITTER2
Sterilizer ID: V2313228
08/28/2023 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.9 F 0.0
2:00 74.7 F 0.0
4:00 93.7 F 0.0
6:00 98.8 F 0.0
8:00 130.4 F 0.0
10:00 161.2 F 1.0
12:00 187.5 F 3.4
14:00 206.3 F 6.4
16:00 216.8 F 9.7
18:00 227.4 F 13.5
20:00 229.2 F 8.2
22:00 239.7 F 12.6
24:00 237.5 F 9.9
26:00 238.4 F 9.5
28:00 247.4 F 14.1
30:00 258.2 F 20.0
32:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.9
2:00 272.6 F 29.4
3:00 271.7 F 28.8
4:00 272.2 F 28.9

Min 271.6 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 214.5 F 0.0
10:00 219.4 F 0.0
15:00 219.6 F 0.0
20:00 219.7 F 0.0
25:00 222.8 F 0.0
30:00 224.3 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:11:20
CYCLE COMPLETE

08/28/2023 09:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____