

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3719

Name: RITTER 1  
Sterilizer ID: V2178198  
09/05/2023 01:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.1 F 0.0  
2:00 180.7 F 2.3  
4:00 192.9 F 6.5  
6:00 212.0 F 12.4  
8:00 224.2 F 7.1  
10:00 237.5 F 12.6  
12:00 238.2 F 10.3  
14:00 235.9 F 8.6  
16:00 251.2 F 15.9  
18:00 263.5 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.4 F 28.6  
2:00 272.3 F 29.0  
3:00 272.0 F 28.4  
4:00 272.6 F 29.1  
  
Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 210.7 F 0.0  
10:00 226.3 F 0.0  
15:00 229.2 F 0.0  
20:00 228.5 F 0.0  
25:00 225.1 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:00  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

09/05/2023 02:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_