
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000385
7:14 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:45
STERILIZING 5:01
136.0 C 227kPa 5:01
136.1 C 230kPa 5:31
136.4 C 233kPa 6:01
136.4 C 232kPa 6:31
136.4 C 233kPa 7:01
136.4 C 233kPa 7:31
136.3 C 232kPa 8:01
136.4 C 233kPa 8:31
136.4 C 233kPa 8:31
VENTING 8:31
AIR DRYING 9:26
CYCLE COMPLETE 9:31

Digital Signature #
B1DE98F966901E0A
