
Midmark M11 - v1.0.5
Total Cycles: 1075

Name: RITTER2
Sterilizer ID: V2313228
05/05/2022 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.1 F 0.0
2:00 132.8 F 0.0
4:00 150.6 F 1.0
6:00 177.3 F 4.9
8:00 210.4 F 9.7
10:00 225.3 F 8.0
12:00 229.0 F 9.6
14:00 234.7 F 9.3
16:00 242.7 F 12.3
18:00 241.0 F 11.0
20:00 253.7 F 17.5
22:00 264.8 F 24.1
24:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.8 F 29.5
2:00 271.3 F 28.6
3:00 272.3 F 29.6
4:00 272.2 F 29.0

Min 271.1 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 227.5 F 0.0
10:00 214.1 F 0.0
15:00 208.5 F 0.0
20:00 204.4 F 0.0
25:00 163.4 F 0.0
30:00 174.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:00
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

05/05/2022 09:17 AM
ACCEPTED: X
REJECTED:
Operator: U02