
Midmark M11 - v1.0.5
Total Cycles: 4497

Name: RITTER2
Sterilizer ID: V2313228
04/08/2025 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.2 F 0.0
2:00 117.5 F 1.0
4:00 118.7 F 3.2
6:00 140.4 F 6.2
8:00 172.5 F 9.8
10:00 188.0 F 10.2
12:00 218.0 F 5.8
14:00 226.7 F 9.7
16:00 236.6 F 10.7
18:00 237.3 F 10.6
20:00 236.0 F 8.8
22:00 247.0 F 14.4
24:00 257.7 F 20.3
26:00 267.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 271.4 F 29.5
2:00 271.6 F 29.5
3:00 271.8 F 29.6
4:00 271.9 F 29.8

Min 271.2 F 29.4
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 201.2 F 0.0
10:00 197.1 F 0.0
15:00 186.9 F 0.0
20:00 181.8 F 0.0
25:00 180.4 F 0.0
30:00 182.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:21
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:06:18
CYCLE COMPLETE

04/08/2025 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____