
Midmark M11 - v1.0.0
Total Cycles: 5303

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2025 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 74.8 F 0.0
4:00 89.2 F 0.6
6:00 124.8 F 3.0
8:00 152.7 F 5.7
10:00 175.3 F 8.9
12:00 193.0 F 12.4
14:00 203.6 F 3.2
16:00 214.2 F 6.7
18:00 225.7 F 10.5
20:00 228.6 F 7.9
22:00 237.1 F 10.4
24:00 235.3 F 8.0
26:00 246.5 F 13.7
28:00 257.7 F 19.7
30:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.8 F 29.2
4:00 271.9 F 28.5

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 198.7 F 0.0
10:00 191.5 F 0.0
15:00 215.8 F 0.0
20:00 234.8 F 0.0
25:00 228.8 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

01/31/2025 10:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____