

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1471

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2022 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.4 F 0.0  
2:00 164.7 F 0.0  
4:00 166.6 F 1.3  
6:00 190.6 F 5.6  
8:00 212.5 F 11.7  
10:00 227.6 F 8.1  
12:00 238.5 F 11.2  
14:00 243.0 F 12.4  
16:00 242.1 F 11.4  
18:00 256.1 F 18.8  
20:00 268.1 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.0 F 28.6  
2:00 272.7 F 29.1  
3:00 272.5 F 29.4  
4:00 271.7 F 28.4

Min 271.5 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 222.9 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 21:08  
STERILIZING: 4:00  
VENTING: 2:47  
DRYING: 3:31  
TOTAL CYCLE: 00:32:24

09/09/2022 09:56 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_