

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001091  
14:09 12/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:31  
STERILIZING 4:30  
136.5 C 220kPa 4:30  
136.4 C 225kPa 5:00  
136.4 C 225kPa 5:30  
136.0 C 221kPa 6:00  
136.3 C 223kPa 6:30  
136.3 C 224kPa 7:00  
136.4 C 223kPa 7:30  
136.2 C 223kPa 8:00  
136.2 C 223kPa 8:00  
VENTING 8:00  
AIR DRYING 8:56  
CYCLE COMPLETE 12:35

-----  
Digital Signature #  
8B3A20BFBCC3268F  
-----