

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001829  
15:18 11/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:00  
STERILIZING 3:00  
136.0 C 220kPa 3:00  
136.1 C 222kPa 3:30  
136.4 C 225kPa 4:00  
136.4 C 223kPa 4:30  
136.1 C 222kPa 5:00  
136.3 C 224kPa 5:30  
136.3 C 224kPa 6:00  
136.4 C 224kPa 6:30  
136.4 C 224kPa 6:30  
VENTING 6:31  
AIR DRYING 7:20  
CYCLE COMPLETE 37:20  
-----

Digital Signature #  
4C04C89F06006AB7  
-----