

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2053

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2022 12:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.9 F 0.0  
2:00 141.3 F 1.0  
4:00 167.1 F 3.7  
6:00 191.2 F 7.5  
8:00 213.7 F 11.9  
10:00 219.4 F 5.9  
12:00 231.6 F 10.6  
14:00 231.3 F 7.4  
16:00 242.3 F 12.5  
18:00 240.8 F 10.8  
20:00 253.5 F 17.2  
22:00 264.3 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.3 F 29.2  
2:00 271.9 F 28.6  
3:00 272.4 F 29.3  
4:00 272.1 F 28.8

Min 271.6 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 200.9 F 0.0  
10:00 219.2 F 0.0  
15:00 222.0 F 0.0  
20:00 224.8 F 0.0  
25:00 228.0 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:59  
CYCLE COMPLETE

06/24/2022 01:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_