
Midmark M11 - v1.0.5
Total Cycles: 155

Name: RITTER2
Sterilizer ID: V2313228
06/29/2021 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 136.8 F 1.5
4:00 152.4 F 4.0
6:00 174.8 F 7.4
8:00 193.1 F 11.1
10:00 213.7 F 4.4
12:00 221.6 F 8.3
14:00 232.0 F 12.8
16:00 234.9 F 9.9
18:00 238.7 F 9.8
20:00 247.0 F 14.3
22:00 258.0 F 20.3
24:00 268.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.5 F 29.6
2:00 272.4 F 30.0
3:00 271.7 F 29.3
4:00 272.5 F 29.7

Min 271.6 F 29.2
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 231.3 F 0.0
10:00 226.9 F 0.0
15:00 222.2 F 0.0
20:00 221.8 F 0.0
25:00 223.2 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:17
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

06/29/2021 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____