

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5134

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2024 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 72.6 F 0.0  
4:00 95.3 F 0.0  
6:00 132.8 F 0.1  
8:00 163.0 F 0.3  
10:00 188.5 F 0.5  
12:00 207.9 F 1.5  
14:00 223.0 F 5.3  
16:00 231.7 F 8.6  
18:00 241.4 F 12.8  
20:00 237.2 F 9.5  
22:00 238.0 F 9.1  
24:00 245.7 F 12.3  
26:00 245.7 F 12.9  
28:00 258.5 F 19.7  
30:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.8 F 28.8  
2:00 272.1 F 28.2  
3:00 271.9 F 28.3  
4:00 272.9 F 28.8

Min 271.6 F 27.9  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 202.5 F 0.0  
10:00 219.1 F 0.0  
15:00 233.1 F 0.0  
20:00 224.2 F 0.0  
25:00 220.6 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 31:06  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:09:12  
CYCLE COMPLETE

11/27/2024 10:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_