
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004372
12:30 15/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:08
STERILIZING 3:11
136.4 C 221kPa 3:11
136.0 C 222kPa 3:41
136.2 C 223kPa 4:11
136.3 C 223kPa 4:41
136.3 C 221kPa 5:11
136.1 C 222kPa 5:41
136.4 C 225kPa 6:11
136.4 C 223kPa 6:41
136.4 C 223kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 7:46

Digital Signature #
C7DA7A02D009037F
