
Midmark M11 - v1.0.5
Total Cycles: 3112

Name: RITTER2
Sterilizer ID: V2313228
11/10/2023 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 189.6 F 3.4
4:00 201.9 F 8.8
6:00 217.7 F 9.5
8:00 229.8 F 9.5
10:00 235.5 F 9.4
12:00 243.9 F 13.5
14:00 244.6 F 13.1
16:00 257.7 F 20.0
18:00 269.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.2 F 29.5
2:00 271.3 F 28.7
3:00 272.2 F 29.7
4:00 272.0 F 29.2

Min 271.1 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 218.1 F 0.0
10:00 205.6 F 0.0
15:00 204.7 F 0.0
20:00 202.9 F 0.0
25:00 200.9 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

11/10/2023 01:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____