

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4343

Name: RITTER 1  
Sterilizer ID: V2178198  
02/28/2024 04:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.3 F 0.0  
2:00 98.4 F 0.8  
4:00 115.4 F 2.8  
6:00 145.1 F 5.7  
8:00 178.1 F 9.2  
10:00 197.9 F 13.5  
12:00 216.3 F 4.9  
14:00 228.6 F 8.6  
16:00 238.2 F 12.7  
18:00 234.9 F 9.1  
20:00 240.5 F 10.8  
22:00 243.4 F 12.4  
24:00 255.0 F 18.3  
26:00 265.4 F 24.6  
28:00 271.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.2  
2:00 272.5 F 29.6  
3:00 271.5 F 29.0  
4:00 272.5 F 29.4

Min 271.5 F 28.9  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 224.2 F 0.0  
10:00 230.0 F 0.0  
15:00 220.6 F 0.0  
20:00 213.6 F 0.0  
25:00 223.9 F 0.0  
30:00 198.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 28:35  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:07:06  
CYCLE COMPLETE

02/28/2024 05:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_