

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006425  
9:27 10/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:08  
STERILIZING 4:05  
136.0 C 220kPa 4:05  
136.2 C 223kPa 4:35  
136.1 C 223kPa 5:05  
136.4 C 225kPa 5:35  
136.4 C 224kPa 6:05  
136.2 C 218kPa 6:35  
136.4 C 224kPa 7:05  
136.4 C 222kPa 7:35  
136.4 C 222kPa 7:35  
VENTING 7:35  
AIR DRYING 8:28  
CYCLE COMPLETE 9:35

-----  
Digital Signature #  
353F23BC9DA82002  
-----