

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 002348  
13:48 03/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:23  
STERILIZING 3:23  
136.0 C 220kPa 3:23  
136.3 C 225kPa 3:53  
136.1 C 224kPa 4:23  
136.4 C 226kPa 4:53  
136.4 C 226kPa 5:23  
136.4 C 223kPa 5:53  
136.3 C 225kPa 6:23  
136.3 C 225kPa 6:53  
136.3 C 225kPa 6:53  
VENTING 6:53  
AIR DRYING 7:46  
CYCLE COMPLETE 16:04  
-----

Digital Signature #  
2B0A4C79B924E51D  
-----