

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3217

Name: RITTER2  
Sterilizer ID: V2313228  
12/14/2023 02:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.2 F 0.0  
2:00 146.1 F 1.7  
4:00 162.6 F 5.1  
6:00 186.1 F 9.8  
8:00 215.9 F 6.4  
10:00 227.8 F 8.7  
12:00 238.8 F 12.3  
14:00 238.8 F 10.8  
16:00 236.7 F 9.2  
18:00 251.0 F 16.1  
20:00 263.0 F 23.0  
22:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.4 F 29.6  
2:00 272.1 F 29.0  
3:00 271.1 F 28.8  
4:00 272.5 F 29.7

Min 271.1 F 28.6  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 226.2 F 0.0  
10:00 226.6 F 0.0  
15:00 188.5 F 0.0  
20:00 181.6 F 0.0  
25:00 187.5 F 0.0  
30:00 207.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:15  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

12/14/2023 03:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_