
Midmark M11 - v1.0.5
Total Cycles: 2643

Name: RITTER2
Sterilizer ID: V2313228
07/13/2023 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 185.9 F 3.7
4:00 204.0 F 9.1
6:00 222.7 F 8.1
8:00 233.5 F 10.9
10:00 233.7 F 8.8
12:00 238.6 F 9.9
14:00 248.1 F 14.9
16:00 260.7 F 22.0
18:00 272.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.8 F 29.9
2:00 271.9 F 29.2
3:00 271.6 F 29.5
4:00 272.8 F 29.8

Min 271.2 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 221.9 F 0.0
10:00 232.5 F 0.0
15:00 225.6 F 0.0
20:00 220.0 F 0.0
25:00 223.5 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

07/13/2023 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____