
Midmark M11 - v1.0.5
Total Cycles: 2640

Name: RITTER2
Sterilizer ID: V2313228
07/13/2023 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.5 F 0.0
4:00 76.0 F 0.0
6:00 98.5 F 0.0
8:00 154.3 F 0.6
10:00 168.1 F 3.8
12:00 198.0 F 7.2
14:00 215.5 F 11.2
16:00 216.3 F 5.2
18:00 224.6 F 8.7
20:00 234.4 F 12.9
22:00 235.2 F 9.6
24:00 237.7 F 9.8
26:00 245.6 F 13.5
28:00 257.1 F 19.5
30:00 267.0 F 25.8
32:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.5 F 29.4
2:00 272.5 F 29.9
3:00 271.5 F 29.1
4:00 272.8 F 29.8

Min 271.4 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 229.9 F 0.0
10:00 218.6 F 0.0
15:00 184.0 F 0.0
20:00 184.1 F 0.0
25:00 192.6 F 0.0
30:00 196.3 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:10:14
CYCLE COMPLETE

07/13/2023 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____