

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000488  
12:12 09/04/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:17  
PRESSURIZING 3:24  
STERILIZING 4:48  
136.0 C 227kPa 4:48  
136.2 C 230kPa 5:18  
136.2 C 231kPa 5:48  
136.3 C 231kPa 6:18  
136.1 C 230kPa 6:48  
136.4 C 233kPa 7:18  
136.3 C 232kPa 7:48  
136.2 C 230kPa 8:18  
136.2 C 230kPa 8:18  
VENTING 8:18  
CYCLE ABORTED 8:38  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:38  
-----

Digital Signature #  
7FFC2CF367D804A7  
-----