
Midmark M11 - v1.0.5
Total Cycles: 2141

Name: RITTER2
Sterilizer ID: V2313228
03/13/2023 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 138.8 F 2.1
4:00 160.7 F 5.2
6:00 185.9 F 9.5
8:00 209.2 F 7.1
10:00 221.1 F 8.3
12:00 233.5 F 12.3
14:00 237.4 F 11.0
16:00 236.7 F 9.6
18:00 250.5 F 16.3
20:00 262.2 F 23.0
22:00 272.9 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 30.0
2:00 271.8 F 29.2
3:00 271.9 F 29.8
4:00 272.5 F 29.9

Min 271.4 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 220.4 F 0.0
10:00 225.0 F 0.0
15:00 228.6 F 0.0
20:00 227.6 F 0.0
25:00 223.4 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

03/13/2023 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____