

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4009

Name: RITTER2  
Sterilizer ID: V2313228  
10/08/2024 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 75.3 F 0.0  
4:00 101.4 F 0.0  
6:00 116.5 F 0.1  
8:00 145.4 F 0.3  
10:00 174.0 F 1.5  
12:00 195.7 F 4.6  
14:00 207.2 F 8.3  
16:00 219.9 F 12.2  
18:00 223.0 F 6.7  
20:00 234.9 F 11.3  
22:00 232.8 F 8.3  
24:00 242.6 F 11.7  
26:00 244.9 F 13.0  
28:00 256.7 F 19.3  
30:00 267.4 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.8 F 29.4  
2:00 271.6 F 28.9  
3:00 272.7 F 29.4  
4:00 271.6 F 28.7  
  
Min 271.5 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 195.8 F 0.0  
10:00 207.2 F 0.0  
15:00 203.4 F 0.0  
20:00 203.9 F 0.0  
25:00 197.5 F 0.0  
30:00 198.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:27  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:38  
CYCLE COMPLETE

10/08/2024 09:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_