
Midmark M11 - v1.0.5
Total Cycles: 1495

Name: RITTER2
Sterilizer ID: V2313228
09/15/2022 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 196.2 F 3.0
4:00 204.2 F 8.5
6:00 224.2 F 8.8
8:00 231.1 F 10.2
10:00 236.0 F 8.0
12:00 242.6 F 11.9
14:00 245.5 F 13.4
16:00 258.3 F 20.5
18:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.4
2:00 271.5 F 29.1
3:00 272.6 F 29.4
4:00 271.4 F 28.9

Min 271.4 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 214.2 F 0.0
10:00 231.9 F 0.0
15:00 228.8 F 0.0
20:00 225.8 F 0.0
25:00 226.8 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 2:45
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

09/15/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____