
Midmark M11 - v1.0.0
Total Cycles: 2252

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 184.5 F 0.4
4:00 202.6 F 2.7
6:00 220.0 F 7.8
8:00 233.8 F 13.2
10:00 237.4 F 10.3
12:00 235.3 F 8.0
14:00 240.4 F 10.0
16:00 249.8 F 14.9
18:00 262.5 F 22.1
20:00 272.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.2 F 28.4
2:00 272.5 F 28.8
3:00 271.7 F 28.0
4:00 272.7 F 28.8

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 227.6 F 0.0
10:00 248.3 F 0.0
15:00 247.7 F 0.0
20:00 231.9 F 0.0
25:00 221.4 F 0.0
Dry Cycle Aborted

FILLING: 1:28
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 28:43
TOTAL CYCLE: 00:57:32

08/25/2022 05:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____