

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007514  
10:23 07/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:54  
STERILIZING 2:47  
136.0 C 220kPa 2:47  
136.1 C 222kPa 3:17  
136.0 C 217kPa 3:47  
136.2 C 221kPa 4:17  
136.4 C 225kPa 4:47  
136.2 C 222kPa 5:17  
136.4 C 222kPa 5:47  
136.3 C 222kPa 6:17  
136.3 C 222kPa 6:17  
VENTING 6:17  
AIR DRYING 7:12  
CYCLE COMPLETE 37:12

-----

Digital Signature #  
92D76B4026AAA535  
-----