
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000334
10:21 08/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:37
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:37

Digital Signature #
2AB7912E4616ACC4
