
Midmark M11 - v1.0.5
Total Cycles: 2600

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.0 F 0.0
2:00 193.6 F 3.3
4:00 196.0 F 8.1
6:00 215.3 F 13.6
8:00 228.6 F 9.2
10:00 235.7 F 9.7
12:00 242.1 F 12.7
14:00 241.4 F 11.5
16:00 254.3 F 18.1
18:00 265.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.7 F 29.8
2:00 271.9 F 29.2
3:00 271.2 F 29.3
4:00 272.9 F 30.3

Min 271.0 F 28.9
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 222.5 F 0.0
10:00 239.6 F 0.0
15:00 236.7 F 0.0
20:00 225.9 F 0.0
25:00 221.8 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

07/04/2023 11:03 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____