
Midmark M11 - v1.0.5
Total Cycles: 2787

Name: RITTER2
Sterilizer ID: V2313228
08/17/2023 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 187.1 F 0.0
4:00 214.0 F 2.8
6:00 229.1 F 8.0
8:00 241.1 F 13.5
10:00 240.8 F 11.6
12:00 240.3 F 10.9
14:00 240.3 F 10.8
16:00 255.0 F 18.3
18:00 267.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.1
2:00 272.2 F 29.3
3:00 272.3 F 28.9
4:00 272.3 F 29.5

Min 271.6 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 225.5 F 0.0
10:00 218.3 F 0.0
15:00 213.5 F 0.0
20:00 191.6 F 0.0
25:00 192.2 F 0.0
30:00 186.6 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

08/17/2023 04:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____