
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001057
7:09 30/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:22
STERILIZING 4:20
136.0 C 220kPa 4:20
136.4 C 220kPa 4:50
136.3 C 222kPa 5:20
136.2 C 220kPa 5:50
136.4 C 223kPa 6:20
136.1 C 221kPa 6:50
136.1 C 221kPa 7:20
136.3 C 221kPa 7:50
136.3 C 221kPa 7:50
VENTING 7:50
AIR DRYING 8:43
CYCLE COMPLETE 8:46

Digital Signature #
3C49AB9B538516AE
