

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002252  
12:42 13/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:01  
STERILIZING 3:03  
136.0 C 220kPa 3:03  
136.3 C 222kPa 3:33  
136.2 C 224kPa 4:03  
136.4 C 224kPa 4:33  
136.4 C 225kPa 5:03  
136.3 C 224kPa 5:33  
136.4 C 224kPa 6:03  
136.3 C 225kPa 6:33  
136.3 C 225kPa 6:33  
VENTING 6:33  
AIR DRYING 7:26  
CYCLE COMPLETE 37:26  
-----

Digital Signature #  
D270B37A11500531  
-----