
Midmark M11 - v1.0.5
Total Cycles: 4380

Name: RITTER2
Sterilizer ID: V2313228
02/26/2025 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 71.3 F 0.2
4:00 76.7 F 0.9
6:00 112.4 F 3.0
8:00 150.2 F 6.1
10:00 165.1 F 9.8
12:00 188.0 F 13.5
14:00 203.8 F 5.0
16:00 216.7 F 8.4
18:00 232.1 F 12.2
20:00 229.5 F 7.7
22:00 240.6 F 12.7
24:00 239.2 F 10.6
26:00 251.3 F 16.6
28:00 261.9 F 22.8
30:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.9 F 29.8
2:00 271.6 F 29.1
3:00 273.0 F 29.9
4:00 271.8 F 29.0

Min 271.5 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 182.6 F 0.0
10:00 196.1 F 0.0
15:00 201.1 F 0.0
20:00 195.1 F 0.0
25:00 193.5 F 0.0
30:00 194.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 30:32
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

02/26/2025 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____