
Midmark M11 - v1.0.0
Total Cycles: 4326

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 166.5 F 1.8
4:00 171.9 F 5.5
6:00 193.7 F 9.9
8:00 213.8 F 5.3
10:00 227.9 F 8.2
12:00 238.3 F 13.2
14:00 238.6 F 10.2
16:00 235.7 F 8.3
18:00 249.2 F 14.9
20:00 261.0 F 21.5
22:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 28.9
2:00 271.5 F 28.7
3:00 272.3 F 28.8
4:00 271.3 F 28.6

Min 271.1 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 219.6 F 0.0
10:00 201.7 F 0.0
15:00 228.1 F 0.0
20:00 214.3 F 0.0
25:00 218.7 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

02/23/2024 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____