
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006296
14:14 12/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:36
STERILIZING 3:37
136.0 C 220kPa 3:37
136.4 C 224kPa 4:07
136.4 C 224kPa 4:37
136.3 C 221kPa 5:07
136.4 C 223kPa 5:37
136.4 C 224kPa 6:07
136.4 C 223kPa 6:37
136.4 C 223kPa 7:07
136.4 C 223kPa 7:07
VENTING 7:07
AIR DRYING 7:59
CYCLE COMPLETE 8:14

Digital Signature #
4D0457B344E1AD07
