
Midmark M11 - v1.0.5
Total Cycles: 645

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.9 F 0.0
2:00 174.5 F 2.7
4:00 196.8 F 7.3
6:00 215.0 F 12.6
8:00 227.1 F 7.9
10:00 239.9 F 13.3
12:00 240.1 F 11.6
14:00 240.0 F 10.9
16:00 253.6 F 17.8
18:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 29.7
2:00 271.4 F 29.2
3:00 272.5 F 29.8
4:00 271.6 F 29.1

Min 271.3 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 232.6 F 0.0
10:00 209.5 F 0.0
15:00 191.2 F 0.0
20:00 208.5 F 0.0
25:00 201.4 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:14
DRYING: 25:51
TOTAL CYCLE: 00:53:59

12/16/2021 02:40 PM
ACCEPTED: X
REJECTED:
Operator: U02