

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006228  
9:24 05/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:29  
STERILIZING 3:36  
136.4 C 223kPa 3:36  
136.4 C 223kPa 4:06  
136.0 C 222kPa 4:36  
136.3 C 221kPa 5:06  
136.4 C 225kPa 5:36  
136.3 C 222kPa 6:06  
136.3 C 223kPa 6:36  
136.3 C 224kPa 7:06  
136.3 C 224kPa 7:06  
VENTING 7:07  
AIR DRYING 7:56  
CYCLE COMPLETE 8:00

-----  
Digital Signature #  
9046DCBC4CEB0D30  
-----