
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000284
11:22 02/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:20
136.0 C 229kPa 4:20
136.4 C 231kPa 4:50
136.3 C 231kPa 5:20
136.4 C 229kPa 5:50
136.3 C 232kPa 6:20
136.2 C 232kPa 6:50
136.2 C 231kPa 7:20
136.4 C 232kPa 7:50
136.4 C 232kPa 7:50
VENTING 7:50
AIR DRYING 8:43
CYCLE COMPLETE 8:45

Digital Signature #
BFB9CE29260C7D13
