
Midmark M11 - v1.0.5
Total Cycles: 2482

Name: RITTER2
Sterilizer ID: V2313228
06/07/2023 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 184.8 F 4.0
4:00 193.8 F 8.9
6:00 214.7 F 8.1
8:00 226.7 F 9.8
10:00 232.4 F 8.5
12:00 243.0 F 13.5
14:00 243.4 F 12.6
16:00 256.4 F 19.5
18:00 267.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.0
2:00 271.9 F 29.3
3:00 271.7 F 29.7
4:00 272.6 F 29.9

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 229.5 F 0.0
10:00 243.2 F 0.0
15:00 234.6 F 0.0
20:00 228.3 F 0.0
25:00 229.3 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

06/07/2023 03:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____