

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004969  
15:41 19/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:27  
PRESSURIZING 1:30  
STERILIZING 2:29  
136.0 C 220kPa 2:29  
136.4 C 222kPa 2:59  
136.4 C 222kPa 3:29  
136.4 C 222kPa 3:59  
136.4 C 223kPa 4:29  
136.1 C 221kPa 4:59  
136.4 C 218kPa 5:29  
136.4 C 222kPa 5:59  
136.4 C 222kPa 5:59  
VENTING 5:59  
AIR DRYING 6:46  
CYCLE COMPLETE 10:16  
-----

Digital Signature #  
B660530DC3EEC919  
-----